

Power-LOAD™ Cot Fastener System

stryker®



EMS Equipment

3800 E. Centre Ave.
Portage, MI 49002 U.S.A.
t: 269 389 2100 f: 269 329 2315
toll free: 800 784 4336

www.ems.stryker.com

Standard Features

- Lifting arms
- Head-end lock LED indicators
- Control panel
- Cot release handles
- Foot-end lock LED indicator
- Linear transfer system
- Inductive charging
- Operation guide

Optional Features

- External power cord
- Wheel guide
- Mass casualty floor mount assembly
- Mass casualty wall mount assembly

Specifications

Model Number	6390
Length	
Overall Length	95 in (241 cm)
Minimum Length	85 in (216 cm)
Width	24.5 in (62 cm)
Weight	
Total Weight	211.5 lb (96.5 kg)
Floor Plate Assembly	16.5 lb (7.5 kg)
Anchor Assembly	23 lb (10.5 kg)
Transfer Assembly	67 lb (30.5 kg)
Trolley Assembly	105 lb (48 kg)
Maximum Weight Capacity*	700 lb (318 kg)
Minimum Operator Required	
Occupied Cot	2
Unoccupied Cot	1
Recommended Loading Height	22 in to 36 in (56 cm to 91 cm)
Battery	12V, 5 Ah Lead Acid Battery (6390-001-468)

Warranty

- One-year parts, labor, and travel

Extended warranties available.

7-year service life.

*Maximum weight capacity represents patient weight. Safe working load of 870 lb (395 kg) represents the sum of the cot total weight and patient.

Stryker reserves the right to change specifications without notice.

In-service video included with every order.

The Power-LOAD cot fastener system is designed to conform to the Federal Specification for the Star-of-Life Ambulance KKK-A-1822.

The yellow and black color scheme is a proprietary trademark of the Stryker Corporation.

Patents pending.

Power-LOAD is currently not available, specifications subject to change without notice. Pending 510(K) clearance, not available for sale. The device has not been authorized as required by the rules of the Federal Communications Commission. The device is not, and may not be, offered for sale or lease, or sold or leased, until authorization is obtained.

Planned Certifications

IPX6 IEC-60601-1 BS EN-1789
AS/NZS-4535 KKK-A-1822

Copyright © 2011 Stryker
DDM/PDF 0411, 10-SEM-3284
Mkt Lit-539 Rev A